

Eviva® EV-500

polysulfone

Eviva® EV-500 polysulfone (PSU) is offered for use in implantable medical devices. Eviva® EV-500 is the neat version of the polymer and Eviva® EV-501 is an opaque white grade.

The manufacturing facility for Eviva® PSU in Alpharetta, GA is ISO 13485 registered and the

relevant aspects of current Good Manufacturing Practices are also applied. Our biomaterial manufacturing processes are carefully validated, and enhanced controls provide product traceability. In addition, all materials are tested in a lab that is ISO 17025 accredited.

An FDA MAF is available.

General

)		
North America	1	
 Medical/Healt 	 Medical/Healthcare Applications 	
• ISO 13485		
Typical Value Unit	Test method	
1.22 to 1.26	ASTM D792	
5.0 to 9.0 g/10 min	ASTM D1238	
Typical Value Unit	Test method	
> 2300 MPa	ASTM D638	
> 70.0 MPa	ASTM D638	
> 20 %	ASTM D638	
Typical Value Unit	Test method	
	ASTM D648	
> 168 °C		
Typical Value Unit	Test method	
Pass	ISO 10993:5	
Pass	ISO 10993:18	
	North America Medical/Healt ISO 13485 Typical Value Unit 1.22 to 1.26 5.0 to 9.0 g/10 min Typical Value Unit > 2300 MPa > 70.0 MPa > 20 % Typical Value Unit > 168 °C Typical Value Unit Pass	

Notes

Typical properties: these are not to be construed as specifications.

www.syensqo.com

Safety Data Sheets (SDS) are available by emailing us or contacting your sales representative. Always consult the appropriate SDS before using any of our products.

Neither Syensqo nor any of its affiliates makes any warranty, express or implied, including merchantability or fitness for use, or accepts any liability in connection with this product, related information or its use. Some applications of which Syensqo's products may be proposed to be used are regulated or restricted by applicable laws and regulations or by national or international standards and in some cases by Syensqo's recommendation, including applications of food/feed, water treatment, medical, pharmaceuticals, and personal care. Only products designated as part of the Solviva® family of biomaterials may be considered as candidates for use in implantable medical devices. The user alone must finally determine suitability of any information or products for any contemplated use in compliance with applicable law, the manner of use and whether any patents are infringed. The information and the products are for use by technically skilled persons at their own discretion and risk and does not relate to the use of this product in combination with any other substance or any other process. This is not a license under any patent or other proprietary right.

All trademarks and registered trademarks are property of the companies that comprise the Syensqo or their respective owners.

© 2025 2023 Syensqo. All rights reserved.

